

Certificate of Analysis

Product	MEDONIC M51-L1 LYSE (200 mL)
Product No	1504511
Lot No	20181127
Manufacturing Date	2018-11-27
Expiration Date	2020-11-26

Test	Specification
pH	2,0 – 2,4
CBC Performance test*	Passes test
Background WBC (x 10 ⁹ /L)	≤ 0,2


*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2019-03-07



Lubna Elabbas
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

