

# Certificate of Analysis



|                           |                              |
|---------------------------|------------------------------|
| <b>Product</b>            | MEDONIC M51-L2 LYSE (500 mL) |
| <b>Product No</b>         | 1504512                      |
| <b>Lot No</b>             | 20190222                     |
| <b>Manufacturing Date</b> | 2019-02-22                   |
| <b>Expiration Date</b>    | 2021-02-21                   |

| <b>Test</b>                           | <b>Specification</b> |
|---------------------------------------|----------------------|
| pH                                    | 4,0 – 4,4            |
| CBC Performance test*                 | Passes test          |
| Background WBC (x 10 <sup>9</sup> /L) | ≤ 0,2                |

\*Performance testing at a reference laboratory. WBC, HGB and WBC scattergrams tested and approved.

Actual values meet specification.

Date: 2019-06-27



Jennifer Abrahamsson  
Quality Assurance

This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

