

# Certificate of Analysis

<b>Product</b>	MEDONIC M51-D DILUENT (20L)
<b>Product No</b>	1504510
<b>Lot No</b>	20190305
<b>Manufacturing Date</b>	2019-03-05
<b>Expiration Date</b>	2021-03-04


<b>Test</b>	<b>Specification</b>
pH	7,4 – 7,8
Osmolality (mOsm/kg)	300 – 320
Conductivity (mS/cm)	15,5 – 16,5
Particles PLT/RBC channel (x 10 <sup>9</sup> /L)	≤ 5

Actual values meet specification.

Date: 2019-05-22



Lubna Elabbas  
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

