

# Certificate of Analysis

**Product** BOULE EASYCLEANER (50 mL)  
**Product No** 1504513  
**Lot No** 2019061601  
**Manufacturing date** 2019-06-16  
**Expiration Date** 2021-06-15

---

Test Hypochlorite	Specification
Hypochlorite concentration	3.50 – 4.95 %
pH	12.60 – 13.20
Blank	PLT $\leq 10 \times 10^9/L$ RBC $\leq 0.02 \times 10^{12}/L$

---

Actual values meet specifications.

Date: 2020-03-05



Lubna Elabbas  
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.  
All information in this document is property of Boule Medical AB.

